

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011223\
 Data File : VX033706.D
 Acq On : 12 Jan 2023 12:26
 Operator : JC/MD
 Sample : 01130-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 230111038-07-TRIP-BLANK

Quant Time: Jan 13 00:01:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jan 07 06:16:21 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/13/2023
 Supervised By : Mahesh Dadoda 01/13/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	11633	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	77480	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	61620	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	29775	31.611	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 105.367%	
60) 4-Bromofluorobenzene	11.079	95	42789	38.528	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 128.433%#	
63) Toluene-d8	8.647	98	80790	40.967	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 136.567%#	
Target Compounds						Qvalue
15) Acetone	2.404	58	16581	119.418	ug/l	84
25) Chloroform	5.105	83	2224m	0.979	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011223\
Data File : VX033706.D
Acq On : 12 Jan 2023 12:26
Operator : JC/MD
Sample : 01130-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
23011038-07-TRIP-BLANK

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/13/2023
Supervised By : Mahesh Dadoda 01/13/2023

Quant Time: Jan 13 00:01:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jan 07 06:16:21 2023
Response via : Initial Calibration

