



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011223\
 Data File : VX033699.D
 Acq On : 12 Jan 2023 08:20
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Sat Jan 07 06:16:21 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	9045	PASS
75	95	30	60	49.5	24253	PASS
95	95	100	100	100.0	49013	PASS
96	95	5	9	6.6	3231	PASS
173	174	0.00	2	1.2	521	PASS
174	95	50	100	87.8	43035	PASS
175	174	5	9	7.7	3297	PASS
176	174	95	101	95.5	41088	PASS
177	176	5	9	7.0	2871	PASS