

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011224\
 Data File : VX039726.D
 Acq On : 12 Jan 2024 10:56
 Operator : JC/MD
 Sample : VX0112MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0112MBL01

Quant Time: Jan 12 22:33:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	86317	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162000	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161533	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72450	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	65116	45.034	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.060%
35) Dibromofluoromethane	5.379	113	51142	46.870	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.740%
50) Toluene-d8	8.647	98	201385	48.936	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.880%
62) 4-Bromofluorobenzene	11.079	95	81699	50.767	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011224\
Data File : VX039726.D
Acq On : 12 Jan 2024 10:56
Operator : JC/MD
Sample : VX0112MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0112MBL01

Quant Time: Jan 12 22:33:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

