

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011224\
 Data File : VX039727.D
 Acq On : 12 Jan 2024 11:19
 Operator : JC/MD
 Sample : VX0112WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0112WBL01

Quant Time: Jan 12 22:33:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	93451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	179208	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172545	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70073	44.762	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.520%
35) Dibromofluoromethane	5.379	113	56846	47.095	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.200%
50) Toluene-d8	8.647	98	221666	48.692	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.380%
62) 4-Bromofluorobenzene	11.079	95	86078	48.352	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.700%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011224\
Data File : VX039727.D
Acq On : 12 Jan 2024 11:19
Operator : JC/MD
Sample : VX0112WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0112WBL01

Quant Time: Jan 12 22:33:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

