

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011224\
 Data File : VX039732.D
 Acq On : 12 Jan 2024 13:21
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 12 22:37:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	165858	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68759	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64834	44.196	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.400%
35) Dibromofluoromethane	5.385	113	53550	47.936	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.880%
50) Toluene-d8	8.647	98	204015	48.422	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.840%
62) 4-Bromofluorobenzene	11.079	95	79151	48.040	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011224\
Data File : VX039732.D
Acq On : 12 Jan 2024 13:21
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 12 22:37:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

