

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011224\
 Data File : VX039742.D
 Acq On : 12 Jan 2024 17:08
 Operator : JC/MD
 Sample : P1118-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-1

Quant Time: Jan 12 22:41:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	98606	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	186075	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182994	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78715	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75550	45.738	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.480%
35) Dibromofluoromethane	5.379	113	59320	47.331	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.660%
50) Toluene-d8	8.647	98	230741	48.815	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.620%
62) 4-Bromofluorobenzene	11.079	95	90038	48.710	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.420%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	20100	21.637	ug/l	90
18) Methyl Acetate	2.703	43	2214	1.139	ug/l	95
20) Methylene Chloride	2.782	84	4340	3.089	ug/l #	89
43) Isopropyl Acetate	6.336	43	91644	24.005	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011224\
Data File : VX039742.D
Acq On : 12 Jan 2024 17:08
Operator : JC/MD
Sample : P1118-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-1

Quant Time: Jan 12 22:41:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

