

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011224\
 Data File : VX039745.D
 Acq On : 12 Jan 2024 18:17
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 12 22:43:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	97111	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	180204	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71826	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73049	44.905	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.800%
35) Dibromofluoromethane	5.385	113	57292	47.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.400%
50) Toluene-d8	8.647	98	220851	48.245	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.480%
62) 4-Bromofluorobenzene	11.079	95	84806	47.375	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.740%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011224\
Data File : VX039745.D
Acq On : 12 Jan 2024 18:17
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 12 22:43:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

