

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011320\
 Data File : VX014580.D
 Acq On : 13 Jan 2020 12:38
 Operator : JC/SP
 Sample : L1074-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CIY1-TB010820

Quant Time: Jan 14 15:13:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	489331	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	756543	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	678506	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	313564	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	271176	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
35) Dibromofluoromethane	5.49	113	223516	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
50) Toluene-d8	8.71	98	894720	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	11.13	95	299659	45.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.86%	
Target Compounds						
20) Methylene Chloride	2.85	84	9423	1.715	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011320\
Data File : VX014580.D
Acq On : 13 Jan 2020 12:38
Operator : JC/SP
Sample : L1074-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CIY1-TB010820

Quant Time: Jan 14 15:13:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

