

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011320\
 Data File : VX014584.D
 Acq On : 13 Jan 2020 14:11
 Operator : JC/SP
 Sample : L1014-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-03-011020

Quant Time: Jan 14 13:43:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	490889	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	763195	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	697397	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	318890	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	280170	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
35) Dibromofluoromethane	5.49	113	227028	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
50) Toluene-d8	8.71	98	916569	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
62) 4-Bromofluorobenzene	11.13	95	313642	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
Target Compounds						
16) Acetone	2.43	43	56762	16.760	ug/l	98
20) Methylene Chloride	2.85	84	39642	7.191	ug/l	95
43) Isopropyl Acetate	6.43	43	113462	9.009	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011320\
Data File : VX014584.D
Acq On : 13 Jan 2020 14:11
Operator : JC/SP
Sample : L1014-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-03-011020

Quant Time: Jan 14 13:43:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

