

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011320\
 Data File : VX014590.D
 Acq On : 13 Jan 2020 16:31
 Operator : JC/SP
 Sample : PB126026ZHE#16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB126026ZHE#16

Quant Time: Jan 14 14:46:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	479267	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	745954	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	686089	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	313903	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	270564	50.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.14%	
35) Dibromofluoromethane	5.49	113	224815	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
50) Toluene-d8	8.71	98	891281	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
62) 4-Bromofluorobenzene	11.13	95	305579	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
Target Compounds						
16) Acetone	2.43	43	29203	6.821	ug/l	94
20) Methylene Chloride	2.84	84	11050	2.053	ug/l #	92
43) Isopropyl Acetate	6.43	43	130220	10.579	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011320\
Data File : VX014590.D
Acq On : 13 Jan 2020 16:31
Operator : JC/SP
Sample : PB126026ZHE#16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB126026ZHE#16

Quant Time: Jan 14 14:46:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

