

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011320\
 Data File : VX014591.D
 Acq On : 13 Jan 2020 16:54
 Operator : JC/SP
 Sample : PB126026ZHE#17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB126026ZHE#17

Quant Time: Jan 14 14:53:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	474185	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	742183	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	686171	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	309829	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	268506	50.72	ug/l	0.00
Spiked Amount			Recovery	=	101.44%	
35) Dibromofluoromethane	5.49	113	220162	49.00	ug/l	0.00
Spiked Amount			Recovery	=	98.00%	
50) Toluene-d8	8.71	98	892233	51.18	ug/l	0.00
Spiked Amount			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	11.13	95	304579	47.59	ug/l	0.00
Spiked Amount			Recovery	=	95.18%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	25978	5.704	ug/l	98
20) Methylene Chloride	2.85	84	10461	1.964	ug/l	94
43) Isopropyl Acetate	6.44	43	134610	10.991	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011320\
Data File : VX014591.D
Acq On : 13 Jan 2020 16:54
Operator : JC/SP
Sample : PB126026ZHE#17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB126026ZHE#17

Quant Time: Jan 14 14:53:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

