

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011321\
 Data File : VX020591.D
 Acq On : 13 Jan 2021 14:02
 Operator : JC/MD
 Sample : M1056-01
 Misc : 5.96g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-01-010821

Quant Time: Jan 13 14:23:00 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	178142	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	286197	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	270438	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	130256	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	108689	50.80	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.60%			
35) Dibromofluoromethane	5.464	113	96800	48.81	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 97.62%			
50) Toluene-d8	8.696	98	343322	48.38	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 96.76%			
62) 4-Bromofluorobenzene	11.122	95	121898	48.65	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 97.30%			
Target Compounds						
20) Methylene Chloride	2.831	84	1650	0.82	ug/l	97
84) 1,2,4-Trimethylbenzene	11.793	105	3477	0.47	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011321\
Data File : VX020591.D
Acq On : 13 Jan 2021 14:02
Operator : JC/MD
Sample : M1056-01
Misc : 5.96g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-01-010821

Quant Time: Jan 13 14:23:00 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

