

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011321\
 Data File : VX020592.D
 Acq On : 13 Jan 2021 14:35
 Operator : JC/MD
 Sample : M1083-02DL2 5000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3DL2

Quant Time: Jan 13 15:20:50 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	189600	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	303646	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	290436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	125080	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	116878	51.32	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.64%			
35) Dibromofluoromethane	5.458	113	104636	49.73	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.46%			
50) Toluene-d8	8.695	98	366622	48.70	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 97.40%			
62) 4-Bromofluorobenzene	11.116	95	127005	47.77	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 95.54%			
Target Compounds						
4) Vinyl Chloride	1.392	62	4457	2.25	ug/l	93
27) cis-1,2-Dichloroethene	4.562	96	57354	25.28	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011321\
Data File : VX020592.D
Acq On : 13 Jan 2021 14:35
Operator : JC/MD
Sample : M1083-02DL2 5000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3DL2

Quant Time: Jan 13 15:20:50 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

