

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011321\
 Data File : VX020597.D
 Acq On : 13 Jan 2021 16:37
 Operator : JC/MD
 Sample : M1069-02
 Misc : 5.82g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FK-GF-02-065-AME

Manual Integrations
 APPROVED

MMDadoda
 1/15/2021 10:06:38 AM

Quant Time: Jan 13 17:08:34 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	193844	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	302580	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	290791	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.067	152	136289	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	113268	48.65	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.30%		
35) Dibromofluoromethane	5.464	113	98933	47.19	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	94.38%		
50) Toluene-d8	8.696	98	361943	48.25	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.50%		
62) 4-Bromofluorobenzene	11.122	95	131624	49.68	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	99.36%		
Target Compounds						Qvalue
18) Methyl Acetate	2.764	43	2450	1.19	ug/l	93
20) Methylene Chloride	2.831	84	1202	0.55	ug/l #	76
25) 2-Butanone	4.666	43	3334m	2.65	ug/l	
27) cis-1,2-Dichloroethene	4.556	96	1298	0.56	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011321\
Data File : VX020597.D
Acq On : 13 Jan 2021 16:37
Operator : JC/MD
Sample : M1069-02
Misc : 5.82g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FK-GF-02-065-AME

Manual Integrations
APPROVED

MMDadoda
1/15/2021 10:06:38 AM

Quant Time: Jan 13 17:08:34 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

