

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011321\
 Data File : VX020599.D
 Acq On : 13 Jan 2021 17:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 13 18:06:10 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.493	0.511	-3.7	97	0.00
3 P	Chloromethane	0.456	0.466	-2.2	98	0.00
4 C	Vinyl Chloride	0.523	0.542	-3.6#	99	0.00
5 T	Bromomethane	0.353	0.294	16.7	80	0.00
6 T	Chloroethane	0.304	0.350	-15.1	102	0.00
7 T	Trichlorofluoromethane	0.867	0.882	-1.7	99	0.00
8 T	Diethyl Ether	0.309	0.310	-0.3	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.486	0.493	-1.4	98	0.00
10 T	Methyl Iodide	0.678	0.702	-3.5	95	0.00
11 T	Tert butyl alcohol	0.117	0.099	15.4	88	-0.02
12 CM	1,1-Dichloroethene	0.477	0.486	-1.9#	99	0.00
13 T	Acrolein	0.083	0.075	9.6	88	0.00
14 T	Allyl chloride	0.663	0.648	2.3	96	0.00
15 T	Acrylonitrile	0.255	0.261	-2.4	99	0.00
16 T	Acetone	0.225	0.204	9.3	96	-0.01
17 T	Carbon Disulfide	1.316	1.269	3.6	94	0.00
18 T	Methyl Acetate	0.532	0.533	-0.2	101	0.00
19 T	Methyl tert-butyl Ether	1.548	1.614	-4.3	99	0.00
20 T	Methylene Chloride	0.567	0.555	2.1	100	0.00
21 T	trans-1,2-Dichloroethene	0.536	0.537	-0.2	98	0.00
22 T	Diisopropyl ether	1.339	1.404	-4.9	99	0.00
23 T	Vinyl Acetate	1.143	1.192	-4.3	98	0.00
24 P	1,1-Dichloroethane	0.889	0.888	0.1	99	0.00
25 T	2-Butanone	0.325	0.320	1.5	97	0.00
26 T	2,2-Dichloropropane	0.827	0.785	5.1	95	0.00
27 T	cis-1,2-Dichloroethene	0.598	0.611	-2.2	100	0.00
28 T	Bromochloromethane	0.362	0.357	1.4	99	0.00
29 T	Tetrahydrofuran	0.198	0.207	-4.5	99	0.00
30 C	Chloroform	0.995	0.992	0.3#	99	0.00
31 T	Cyclohexane	0.706	0.738	-4.5	98	0.00
32 T	1,1,1-Trichloroethane	0.895	0.894	0.1	98	0.00
33 S	1,2-Dichloroethane-d4	0.601	0.599	0.3	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.346	0.340	1.7	102	0.00
36 T	1,1-Dichloropropene	0.457	0.459	-0.4	98	0.00
37 T	Ethyl Acetate	0.415	0.412	0.7	98	0.00
38 T	Carbon Tetrachloride	0.533	0.525	1.5	97	0.00
39 T	Methylcyclohexane	0.522	0.540	-3.4	97	0.00
40 TM	Benzene	1.373	1.404	-2.3	99	0.00
41 T	Methacrylonitrile	0.217	0.221	-1.8	96	0.00
42 TM	1,2-Dichloroethane	0.490	0.498	-1.6	100	0.00
43 T	Isopropyl Acetate	0.643	0.653	-1.6	98	0.00
44 TM	Trichloroethene	0.408	0.407	0.2	99	0.00
45 C	1,2-Dichloropropane	0.337	0.343	-1.8#	100	0.00
46 T	Dibromomethane	0.256	0.256	0.0	99	0.00
47 T	Bromodichloromethane	0.522	0.514	1.5	97	0.00
48 T	Methyl methacrylate	0.313	0.331	-5.8	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011321\
 Data File : VX020599.D
 Acq On : 13 Jan 2021 17:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 13 18:06:10 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	99	-0.03
50 S	Toluene-d8	1.240	1.224	1.3	100	0.00
51 T	4-Methyl-2-Pentanone	0.399	0.415	-4.0	97	0.00
52 CM	Toluene	0.873	0.902	-3.3#	99	0.00
53 T	t-1,3-Dichloropropene	0.532	0.535	-0.6	97	0.00
54 T	cis-1,3-Dichloropropene	0.569	0.581	-2.1	98	0.00
55 T	1,1,2-Trichloroethane	0.369	0.375	-1.6	99	0.00
56 T	Ethyl methacrylate	0.478	0.505	-5.6	98	0.00
57 T	1,3-Dichloropropane	0.599	0.608	-1.5	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.226	0.251	-11.1	98	0.00
59 T	2-Hexanone	0.299	0.311	-4.0	96	0.00
60 T	Dibromochloromethane	0.417	0.417	0.0	97	0.00
61 T	1,2-Dibromoethane	0.394	0.399	-1.3	98	0.00
62 S	4-Bromofluorobenzene	0.438	0.450	-2.7	113	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.452	0.433	4.2	98	0.00
65 PM	Chlorobenzene	1.052	1.047	0.5	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.413	-2.7	108	0.00
67 C	Ethyl Benzene	1.732	1.822	-5.2#	107	0.00
68 T	m/p-Xylenes	0.668	0.697	-4.3	100	0.00
69 T	o-Xylene	0.615	0.656	-6.7	104	0.00
70 T	Styrene	1.050	1.146	-9.1	103	0.00
71 P	Bromoform	0.326	0.324	0.6	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
73 T	Isopropylbenzene	3.371	3.374	-0.1	111	0.00
74 T	N-amyl acetate	1.072	1.063	0.8	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.098	1.093	0.5	114	0.00
76 T	1,2,3-Trichloropropane	1.053	1.015	3.6	109	0.00
77 T	Bromobenzene	0.903	0.867	4.0	110	0.00
78 T	n-propylbenzene	3.726	3.862	-3.7	111	0.00
79 T	2-Chlorotoluene	2.299	2.296	0.1	111	0.00
80 T	1,3,5-Trimethylbenzene	2.825	2.950	-4.4	113	0.00
81 T	trans-1,4-Dichloro-2-butene	0.387	0.348	10.1	103	0.00
82 T	4-Chlorotoluene	2.728	2.745	-0.6	112	0.00
83 T	tert-Butylbenzene	2.713	2.731	-0.7	109	0.00
84 T	1,2,4-Trimethylbenzene	2.824	2.970	-5.2	110	0.00
85 T	sec-Butylbenzene	3.166	3.242	-2.4	111	0.00
86 T	p-Isopropyltoluene	2.940	3.039	-3.4	109	0.00
87 T	1,3-Dichlorobenzene	1.651	1.572	4.8	108	0.00
88 T	1,4-Dichlorobenzene	1.692	1.642	3.0	113	0.00
89 T	n-Butylbenzene	2.544	2.601	-2.2	112	0.00
90 T	Hexachloroethane	0.555	0.511	7.9	107	0.00
91 T	1,2-Dichlorobenzene	1.630	1.601	1.8	114	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.258	0.253	1.9	118	0.00
93 T	1,2,4-Trichlorobenzene	1.087	1.107	-1.8	112	0.00
94 T	Hexachlorobutadiene	0.593	0.505	14.8	103	0.00
95 T	Naphthalene	3.166	3.458	-9.2	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011321\
Data File : VX020599.D
Acq On : 13 Jan 2021 17:23
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 13 18:06:10 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.105	1.107	-0.2	111	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6