

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011420\
 Data File : VX014606.D
 Acq On : 14 Jan 2020 16:23
 Operator : JC/SP
 Sample : L1125-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FEQCFB-01-14-20

Quant Time: Jan 15 01:03:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	469465	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	733001	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	667719	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	305264	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	268423	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
35) Dibromofluoromethane	5.48	113	222132	50.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.12%	
50) Toluene-d8	8.71	98	868628	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	11.13	95	292652	46.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.60%	
Target Compounds						
16) Acetone	2.43	43	14771	1.467	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011420\
Data File : VX014606.D
Acq On : 14 Jan 2020 16:23
Operator : JC/SP
Sample : L1125-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FEQCFB-01-14-20

Quant Time: Jan 15 01:03:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

