

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX011420\
 Data File : VX014596.D
 Acq On : 14 Jan 2020 09:38
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 15 00:23:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.557	0.678	-21.7#	91	0.00
3 P	Chloromethane	0.637	0.673	-5.7	89	0.00
4 C	Vinyl Chloride	0.710	0.781	-10.0#	89	0.00
5 T	Bromomethane	0.471	0.460	2.3	91	0.00
6 T	Chloroethane	0.420	0.433	-3.1	92	0.00
7 T	Trichlorofluoromethane	0.806	0.850	-5.5	91	0.00
8 T	Diethyl Ether	0.368	0.369	-0.3	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.490	0.520	-6.1	93	0.00
10 T	Methyl Iodide	0.587	0.684	-16.5	90	0.00
11 T	Tert butyl alcohol	0.117	0.109	6.8	79	0.00
12 CM	1,1-Dichloroethene	0.494	0.505	-2.2#	90	0.00
13 T	Acrolein	0.093	0.083	10.8	76	0.00
14 T	Allyl chloride	0.867	0.912	-5.2	91	0.00
15 T	Acrylonitrile	0.281	0.291	-3.6	90	0.00
16 T	Acetone	0.303	0.326	-7.6	97	0.00
17 T	Carbon Disulfide	1.378	1.349	2.1	87	0.00
18 T	Methyl Acetate	0.802	0.880	-9.7	90	0.00
19 T	Methyl tert-butyl Ether	1.585	1.674	-5.6	90	0.00
20 T	Methylene Chloride	0.562	0.573	-2.0	92	0.00
21 T	trans-1,2-Dichloroethene	0.539	0.545	-1.1	91	0.00
22 T	Diisopropyl ether	1.675	1.844	-10.1	94	0.00
23 T	Vinyl Acetate	1.297	1.439	-10.9	95	0.00
24 P	1,1-Dichloroethane	0.911	0.986	-8.2	92	0.00
25 T	2-Butanone	0.417	0.448	-7.4	92	0.00
26 T	2,2-Dichloropropane	0.772	0.814	-5.4	91	0.00
27 T	cis-1,2-Dichloroethene	0.603	0.619	-2.7	91	0.00
28 T	Bromochloromethane	0.384	0.388	-1.0	91	-0.01
29 T	Tetrahydrofuran	0.247	0.259	-4.9	87	0.00
30 C	Chloroform	0.891	0.965	-8.3#	93	-0.01
31 T	Cyclohexane	0.853	0.939	-10.1	91	0.00
32 T	1,1,1-Trichloroethane	0.762	0.824	-8.1	92	-0.01
33 S	1,2-Dichloroethane-d4	0.558	0.550	1.4	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.303	0.311	-2.6	87	-0.01
36 T	1,1-Dichloropropene	0.470	0.509	-8.3	92	0.00
37 T	Ethyl Acetate	0.511	0.524	-2.5	86	-0.01
38 T	Carbon Tetrachloride	0.433	0.482	-11.3	91	0.00
39 T	Methylcyclohexane	0.591	0.651	-10.2	91	0.00
40 TM	Benzene	1.396	1.527	-9.4	92	0.00
41 T	Methacrylonitrile	0.272	0.302	-11.0	90	0.00
42 TM	1,2-Dichloroethane	0.488	0.523	-7.2	92	0.00
43 T	Isopropyl Acetate	0.825	0.882	-6.9	87	0.00
44 TM	Trichloroethene	0.401	0.427	-6.5	90	0.00
45 C	1,2-Dichloropropane	0.356	0.401	-12.6#	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX011420\
 Data File : VX014596.D
 Acq On : 14 Jan 2020 09:38
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 15 00:23:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.248	0.255	-2.8	90	0.00
47 T	Bromodichloromethane	0.452	0.515	-13.9	93	0.00
48 T	Methyl methacrylate	0.395	0.442	-11.9	89	0.00
49 T	1,4-Dioxane	0.008	0.008	0.0	82	0.00
50 S	Toluene-d8	1.175	1.189	-1.2	86	0.00
51 T	4-Methyl-2-Pentanone	0.503	0.549	-9.1	88	0.00
52 CM	Toluene	0.893	0.966	-8.2#	90	0.00
53 T	t-1,3-Dichloropropene	0.526	0.568	-8.0	89	0.00
54 T	cis-1,3-Dichloropropene	0.568	0.633	-11.4	90	0.00
55 T	1,1,2-Trichloroethane	0.359	0.391	-8.9	92	0.00
56 T	Ethyl methacrylate	0.550	0.614	-11.6	89	0.00
57 T	1,3-Dichloropropane	0.600	0.654	-9.0	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.187	0.225	-20.3#	97	0.00
59 T	2-Hexanone	0.398	0.449	-12.8	91	0.00
60 T	Dibromochloromethane	0.361	0.415	-15.0	92	0.00
61 T	1,2-Dibromoethane	0.380	0.401	-5.5	90	0.00
62 S	4-Bromofluorobenzene	0.431	0.422	2.1	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.501	0.506	-1.0	84	0.00
65 PM	Chlorobenzene	1.085	1.148	-5.8	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.420	-8.5	91	0.00
67 C	Ethyl Benzene	1.849	2.026	-9.6#	91	0.00
68 T	m/p-Xylenes	0.714	0.778	-9.0	91	0.00
69 T	o-Xylene	0.681	0.747	-9.7	91	0.00
70 T	Styrene	1.175	1.304	-11.0	92	0.00
71 P	Bromoform	0.317	0.345	-8.8	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.511	3.834	-9.2	92	0.00
74 T	N-amyl acetate	1.549	1.635	-5.6	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.126	1.204	-6.9	93	0.00
76 T	1,2,3-Trichloropropane	1.035	1.046	-1.1	93	0.00
77 T	Bromobenzene	0.974	0.993	-2.0	91	0.00
78 T	n-propylbenzene	4.021	4.478	-11.4	94	0.00
79 T	2-Chlorotoluene	2.397	2.550	-6.4	92	0.00
80 T	1,3,5-Trimethylbenzene	2.915	3.204	-9.9	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.391	0.417	-6.6	86	0.00
82 T	4-Chlorotoluene	2.813	3.016	-7.2	93	0.00
83 T	tert-Butylbenzene	2.754	3.021	-9.7	94	0.00
84 T	1,2,4-Trimethylbenzene	2.946	3.242	-10.0	93	0.00
85 T	sec-Butylbenzene	3.424	3.799	-11.0	93	0.00
86 T	p-Isopropyltoluene	3.177	3.524	-10.9	93	0.00
87 T	1,3-Dichlorobenzene	1.747	1.789	-2.4	91	0.00
88 T	1,4-Dichlorobenzene	1.792	1.791	0.1	90	0.00
89 T	n-Butylbenzene	2.905	3.132	-7.8	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX011420\
Data File : VX014596.D
Acq On : 14 Jan 2020 09:38
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 15 00:23:29 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.558	0.611	-9.5	91	0.00
91 T	1,2-Dichlorobenzene	1.712	1.776	-3.7	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.253	0.257	-1.6	85	0.00
93 T	1,2,4-Trichlorobenzene	1.310	1.260	3.8	88	0.00
94 T	Hexachlorobutadiene	0.610	0.630	-3.3	93	0.00
95 T	Naphthalene	3.730	3.619	3.0	84	0.00
96 T	1,2,3-Trichlorobenzene	1.278	1.243	2.7	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6