

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
 Data File : VX020602.D
 Acq On : 14 Jan 2021 10:56
 Operator : JC/MD
 Sample : VX0114MBL02
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0114MBL02

Quant Time: Jan 15 05:08:14 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	178436	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	289366	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.091	117	277129	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	122892	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	112420	52.45	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.90%
35) Dibromofluoromethane	5.458	113	99918	49.83	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.66%
50) Toluene-d8	8.695	98	346875	48.35	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.70%
62) 4-Bromofluorobenzene	11.122	95	118855	46.91	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.82%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
Data File : VX020602.D
Acq On : 14 Jan 2021 10:56
Operator : JC/MD
Sample : VX0114MBL02
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0114MBL02

Quant Time: Jan 15 05:08:14 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

