

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
 Data File : VX020613.D
 Acq On : 14 Jan 2021 15:26
 Operator : JC/MD
 Sample : PB134118ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB134118ZHE#02

Quant Time: Jan 15 05:11:01 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	182410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	300151	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	304200	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	132268	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	114399	52.21	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.42%			
35) Dibromofluoromethane	5.464	113	100698	48.42	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 96.84%			
50) Toluene-d8	8.695	98	380679	51.16	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.32%			
62) 4-Bromofluorobenzene	11.122	95	137927	52.48	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 104.96%			
Target Compounds						
16) Acetone	2.416	43	8172	7.09	ug/l	99
20) Methylene Chloride	2.837	84	7940	3.84	ug/l	87
43) Isopropyl Acetate	6.409	43	14821	3.84	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
Data File : VX020613.D
Acq On : 14 Jan 2021 15:26
Operator : JC/MD
Sample : PB134118ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB134118ZHE#02

Quant Time: Jan 15 05:11:01 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

