

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
 Data File : VX020614.D
 Acq On : 14 Jan 2021 15:50
 Operator : JC/MD
 Sample : PB134118ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB134118ZHE#03

Quant Time: Jan 15 05:11:12 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	189785	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	298725	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	312803	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	138917	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	119402	52.38	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.76%	
35) Dibromofluoromethane	5.465	113	106553	51.48	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.96%	
50) Toluene-d8	8.696	98	386885	52.24	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.48%	
62) 4-Bromofluorobenzene	11.122	95	140882	53.86	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.72%	
Target Compounds						
						Qvalue
16) Acetone	2.416	43	7957	6.38	ug/l	96
20) Methylene Chloride	2.831	84	7633	3.54	ug/l	97
43) Isopropyl Acetate	6.409	43	15061	3.92	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
Data File : VX020614.D
Acq On : 14 Jan 2021 15:50
Operator : JC/MD
Sample : PB134118ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB134118ZHE#03

Quant Time: Jan 15 05:11:12 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

