

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
 Data File : VX020615.D
 Acq On : 14 Jan 2021 16:13
 Operator : JC/MD
 Sample : PB134118ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB134118ZHE#04

Quant Time: Jan 15 05:11:23 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	191612	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	320286	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	291834	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	136063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	118013	51.28	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.56%
35) Dibromofluoromethane	5.464	113	105866	47.70	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.40%
50) Toluene-d8	8.695	98	398858	50.23	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.46%
62) 4-Bromofluorobenzene	11.122	95	140415	50.07	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.14%
Target Compounds						
					Qvalue	
16) Acetone	2.416	43	8958	7.57	ug/l	98
20) Methylene Chloride	2.831	84	19761	9.09	ug/l	98
43) Isopropyl Acetate	6.403	43	14485	3.51	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
Data File : VX020615.D
Acq On : 14 Jan 2021 16:13
Operator : JC/MD
Sample : PB134118ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB134118ZHE#04

Quant Time: Jan 15 05:11:23 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

