

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
 Data File : VX020616.D
 Acq On : 14 Jan 2021 16:36
 Operator : JC/MD
 Sample : PB134118ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB134118ZHE#05

Quant Time: Jan 15 05:11:34 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	187981	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	313314	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	303091	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	126063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	115910	51.34	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.68%
35) Dibromofluoromethane	5.458	113	103934	47.88	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.76%
50) Toluene-d8	8.696	98	385751	49.66	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.32%
62) 4-Bromofluorobenzene	11.122	95	136388	49.72	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.44%
Target Compounds						
					Qvalue	
16) Acetone	2.416	43	9151	8.05	ug/l	96
20) Methylene Chloride	2.831	84	20746	9.73	ug/l	96
43) Isopropyl Acetate	6.403	43	14489	3.59	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
Data File : VX020616.D
Acq On : 14 Jan 2021 16:36
Operator : JC/MD
Sample : PB134118ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB134118ZHE#05

Quant Time: Jan 15 05:11:34 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

