

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
 Data File : VX020622.D
 Acq On : 14 Jan 2021 18:54
 Operator : JC/MD
 Sample : M1119-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-BRI-SO-011221

Quant Time: Jan 15 05:12:43 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	185775	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	311178	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	286943	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	135726	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	115718	51.86	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.72%	
35) Dibromofluoromethane	5.465	113	104911	48.66	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.32%	
50) Toluene-d8	8.696	98	389092	50.43	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.86%	
62) 4-Bromofluorobenzene	11.122	95	135672	49.80	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 99.60%	
Target Compounds						Qvalue
16) Acetone	2.416	43	10862	10.47	ug/l	97
20) Methylene Chloride	2.831	84	21119	10.02	ug/l	96
25) 2-Butanone	4.629	43	6030	5.00	ug/l	93
43) Isopropyl Acetate	6.409	43	13235	3.31	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011421\
Data File : VX020622.D
Acq On : 14 Jan 2021 18:54
Operator : JC/MD
Sample : M1119-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-BRI-SO-011221

Quant Time: Jan 15 05:12:43 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

