

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011422\
 Data File : VX026448.D
 Acq On : 14 Jan 2022 12:15
 Operator : JC/MD
 Sample : N1123-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 220112003-05-TRIP-BLANK

Quant Time: Jan 17 05:56:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 10 12:50:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.910	128	16485	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	93250	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	77128	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	40519	30.567	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.900%
60) 4-Bromofluorobenzene	11.085	95	37055	28.787	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	95.967%
63) Toluene-d8	8.653	98	106664	30.104	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.333%
Target Compounds						
25) Chloroform	5.093	83	2253	1.192	ug/l	Qvalue 86
41) Bromodichloromethane	7.824	83	505	0.341	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011422\
Data File : VX026448.D
Acq On : 14 Jan 2022 12:15
Operator : JC/MD
Sample : N1123-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
220112003-05-TRIP-BLANK

Quant Time: Jan 17 05:56:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 10 12:50:55 2022
Response via : Initial Calibration

