


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011422\
 Data File : VX026443.D
 Acq On : 14 Jan 2022 09:33
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Mon Jan 10 12:50:55 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	5630	PASS
75	95	30	60	52.0	14878	PASS
95	95	100	100	100.0	28624	PASS
96	95	5	9	6.4	1832	PASS
173	174	0.00	2	1.3	256	PASS
174	95	50	100	70.4	20141	PASS
175	174	5	9	7.7	1554	PASS
176	174	95	101	98.4	19820	PASS
177	176	5	9	6.4	1266	PASS