

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
 Data File : VX033730.D
 Acq On : 13 Jan 2023 18:15
 Operator : JC/MD
 Sample : 01156-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAGRE

Quant Time: Jan 16 05:38:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	169077	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206357	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220926	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83030	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89652	50.360	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.720%	
35) Dibromofluoromethane	5.385	113	76188	59.459	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 118.920%	
50) Toluene-d8	8.647	98	282660	60.544	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 121.080%#	
62) 4-Bromofluorobenzene	11.079	95	98959	57.090	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 114.180%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
Data File : VX033730.D
Acq On : 13 Jan 2023 18:15
Operator : JC/MD
Sample : 01156-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAGRE

Quant Time: Jan 16 05:38:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

