

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
 Data File : VX033731.D
 Acq On : 13 Jan 2023 18:38
 Operator : JC/MD
 Sample : PB150223ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150223ZHE#03

Quant Time: Jan 16 05:38:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	166768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	249754	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91445	52.078	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.160%
35) Dibromofluoromethane	5.385	113	76811	49.529	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.060%
50) Toluene-d8	8.647	98	217999	38.580	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	77.160%#
62) 4-Bromofluorobenzene	11.079	95	99996	47.664	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.320%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	28357	33.720	ug/l	97
25) 2-Butanone	4.574	43	8363	7.289	ug/l	100
37) Ethyl Acetate	4.714	43	24604	9.052	ug/l	99
43) Isopropyl Acetate	6.336	43	104069	29.125	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
Data File : VX033731.D
Acq On : 13 Jan 2023 18:38
Operator : JC/MD
Sample : PB150223ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150223ZHE#03

Quant Time: Jan 16 05:38:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

