

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
 Data File : VX033733.D
 Acq On : 13 Jan 2023 19:24
 Operator : JC/MD
 Sample : PB150223ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150223ZHE#05

Quant Time: Jan 16 05:39:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	116603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	178786	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	175436	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81650	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71508	58.244	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.480%
35) Dibromofluoromethane	5.385	113	54877	49.431	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.860%
50) Toluene-d8	8.647	98	217156	53.686	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.380%
62) 4-Bromofluorobenzene	11.079	95	79360	52.843	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.680%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	21717	38.177	ug/l	92
25) 2-Butanone	4.580	43	6435	8.022	ug/l	93
37) Ethyl Acetate	4.721	43	18436	9.617	ug/l	98
43) Isopropyl Acetate	6.342	43	81452	31.844	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
Data File : VX033733.D
Acq On : 13 Jan 2023 19:24
Operator : JC/MD
Sample : PB150223ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150223ZHE#05

Quant Time: Jan 16 05:39:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

