

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
 Data File : VX033734.D
 Acq On : 13 Jan 2023 19:47
 Operator : JC/MD
 Sample : PB150223ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150223ZHE#06

Quant Time: Jan 16 05:39:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	126017	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	191058	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84005	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77582	58.471	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 116.940%	
35) Dibromofluoromethane	5.385	113	59445	50.107	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.220%	
50) Toluene-d8	8.647	98	236443	54.700	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 109.400%	
62) 4-Bromofluorobenzene	11.079	95	86822	54.099	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.200%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	22587	36.250	ug/l	97
25) 2-Butanone	4.574	43	6728	7.761	ug/l	99
37) Ethyl Acetate	4.715	43	21090	10.508	ug/l	96
43) Isopropyl Acetate	6.342	43	88344	32.320	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
Data File : VX033734.D
Acq On : 13 Jan 2023 19:47
Operator : JC/MD
Sample : PB150223ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150223ZHE#06

Quant Time: Jan 16 05:39:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

