

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
 Data File : VX033736.D
 Acq On : 13 Jan 2023 20:34
 Operator : JC/MD
 Sample : PB150223ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150223ZHE#08

Quant Time: Jan 16 05:39:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	154205	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238897	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88111	54.268	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	108.540%	
35) Dibromofluoromethane	5.385	113	72879	49.129	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.260%	
50) Toluene-d8	8.647	98	282449	52.258	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	104.520%	
62) 4-Bromofluorobenzene	11.079	95	99906	49.785	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	99.580%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	33339	46.413	ug/l	96
25) 2-Butanone	4.574	43	5676	5.350	ug/l	97
37) Ethyl Acetate	4.721	43	24516	9.556	ug/l	99
43) Isopropyl Acetate	6.336	43	109823	32.132	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
Data File : VX033736.D
Acq On : 13 Jan 2023 20:34
Operator : JC/MD
Sample : PB150223ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150223ZHE#08

Quant Time: Jan 16 05:39:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

