

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
 Data File : VX033737.D
 Acq On : 13 Jan 2023 20:57
 Operator : JC/MD
 Sample : PB150223ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150223ZHE#09

Quant Time: Jan 16 05:39:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124474	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	190857	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240509	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114920	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76006	57.993	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.980%
35) Dibromofluoromethane	5.385	113	58718	49.546	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.100%
50) Toluene-d8	8.647	98	305124	70.663	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	141.320%#
62) 4-Bromofluorobenzene	11.079	95	108281	67.541	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	135.080%#
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	25368	43.004	ug/l	94
25) 2-Butanone	4.580	43	4158	4.856	ug/l	91
37) Ethyl Acetate	4.714	43	18653	8.956	ug/l	97
43) Isopropyl Acetate	6.336	43	85574	31.340	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
Data File : VX033737.D
Acq On : 13 Jan 2023 20:57
Operator : JC/MD
Sample : PB150223ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150223ZHE#09

Quant Time: Jan 16 05:39:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

