

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
 Data File : VX033739.D
 Acq On : 13 Jan 2023 21:43
 Operator : JC/MD
 Sample : PB150223ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150223ZHE#02

Quant Time: Jan 16 05:40:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	114024	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	172391	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	170215	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77489	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69315	57.735	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 115.480%			
35) Dibromofluoromethane	5.385	113	52907	49.425	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.840%			
50) Toluene-d8	8.647	98	210793	54.046	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 108.100%			
62) 4-Bromofluorobenzene	11.079	95	76948	53.138	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 106.280%			
Target Compounds					Qvalue	
16) Acetone	2.392	43	21729	39.364	ug/l	100
25) 2-Butanone	4.574	43	6236	7.950	ug/l	95
37) Ethyl Acetate	4.715	43	18930	10.437	ug/l	98
43) Isopropyl Acetate	6.342	43	82525	33.460	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
Data File : VX033739.D
Acq On : 13 Jan 2023 21:43
Operator : JC/MD
Sample : PB150223ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150223ZHE#02

Quant Time: Jan 16 05:40:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

