


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011423\
Data File : VX033718.D
Acq On : 13 Jan 2023 13:20
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Title : SW846 8260
Last Update : Wed Jan 11 14:48:42 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.1	8346	PASS
75	95	30	60	49.0	21467	PASS
95	95	100	100	100.0	43789	PASS
96	95	5	9	6.8	2999	PASS
173	174	0.00	2	0.8	297	PASS
174	95	50	100	86.5	37877	PASS
175	174	5	9	7.9	3007	PASS
176	174	95	101	97.0	36757	PASS
177	176	5	9	6.6	2414	PASS