

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011519\
 Data File : VX007054.D
 Acq On : 15 Jan 2019 13:20
 Operator : JC/SP
 Sample : K1149-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW002-190114-01

Quant Time: Jan 16 00:28:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	578176	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	888640	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	830083	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	407766	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	312557	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
35) Dibromofluoromethane	5.50	113	278562	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
50) Toluene-d8	8.71	98	1064876	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
62) 4-Bromofluorobenzene	11.13	95	377250	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
Target Compounds						
16) Acetone	2.45	43	15858	5.309	ug/l	99
37) Ethyl Acetate	4.86	43	18918	2.284	ug/l #	94
40) Benzene	6.14	78	18056	0.766	ug/l	95
95) Naphthalene	13.83	128	21868	0.744	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011519\
Data File : VX007054.D
Acq On : 15 Jan 2019 13:20
Operator : JC/SP
Sample : K1149-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW002-190114-01

Quant Time: Jan 16 00:28:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

