

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011519\
 Data File : VX007057.D
 Acq On : 15 Jan 2019 14:40
 Operator : JC/SP
 Sample : K1150-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS038-190114

Quant Time: Jan 16 00:36:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	559218	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	872802	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	811010	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	399131	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	300807	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
35) Dibromofluoromethane	5.51	113	268839	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
50) Toluene-d8	8.71	98	1045778	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
62) 4-Bromofluorobenzene	11.13	95	372774	52.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.52%	
Target Compounds						
16) Acetone	2.45	43	9449	3.270	ug/l #	88
27) cis-1,2-Dichloroethene	4.60	96	36323	5.749	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011519\
Data File : VX007057.D
Acq On : 15 Jan 2019 14:40
Operator : JC/SP
Sample : K1150-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS038-190114

Quant Time: Jan 16 00:36:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

