

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX011519\
 Data File : VX007053.D
 Acq On : 15 Jan 2019 12:54
 Operator : JC/SP
 Sample : K1150-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-190114

Quant Time: Jan 16 09:23:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	592080	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	908727	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	855100	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	413673	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	314622	50.85	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.70%	
35) Dibromofluoromethane	5.51	113	279436	48.02	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.04%	
50) Toluene-d8	8.71	98	1083887	50.88	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.76%	
62) 4-Bromofluorobenzene	11.13	95	392736	52.89	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	105.78%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011519\
Data File : VX007053.D
Acq On : 15 Jan 2019 12:54
Operator : JC/SP
Sample : K1150-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-190114

Quant Time: Jan 16 09:23:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX007053.D

Acq: 15 Jan 2019 12:54

Instrument :

MSVOA_X

ClientSampleId :


TB-01-190114

Tot Ion:168 Resp: 592080

Ion Ratio Lower Upper




168 100

99 48.7 41.8 62.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 50.846 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX007053.D


Acq: 15 Jan 2019 12:54

Tgt Ion: 65 Resp: 314622

Ion Ratio Lower Upper



65 100


67 52.5 0.0 105.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX007053.D

Acq: 15 Jan 2019 12:54

Instrument :

MSVOA_X

ClientSampleId :

TB-01-190114


Tot Ion:114 Resp: 908727

Ion Ratio Lower Upper

114 100

63 18.1 0.0 34.8




88 15.3 0.0 28.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 48.018 ug/l

RT: 5.51 min Scan# 714

Delta R.T. 0.01 min

Lab File: VX007053.D

Acq: 15 Jan 2019 12:54


Tgt Ion:113 Resp: 279436

Ion Ratio Lower Upper

113 100

111 101.6 81.6 122.4



192 21.8 16.7 25.1




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 50.884 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX007053.D
Acq: 15 Jan 2019 12:54




Instrument :
MSVOA_X
ClientSampleId :
TB-01-190114


Tot Ion: 98 Resp: 1083887
Ion Ratio Lower Upper
98 100
100 64.0 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 52.886 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. 0.00 min
Lab File: VX007053.D
Acq: 15 Jan 2019 12:54

Tgt Ion: 95 Resp: 392736
Ion Ratio Lower Upper
95 100
174 91.2 0.0 182.2
176 87.5 0.0 178.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.01 min
Lab File: VX007053.D
Acq: 15 Jan 2019 12:54

Instrument :
MSVOA_X
ClientSampleId :
TB-01-190114

Tot Ion:117 Resp: 855100
Ion Ratio Lower Upper
117 100
82 52.8 41.0 61.6
119 32.1 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX007053.D
Acq: 15 Jan 2019 12:54

Tgt Ion:152 Resp: 413673
Ion Ratio Lower Upper
152 100
115 55.0 39.1 117.2
150 157.1 0.0 344.8

