

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX011520\
 Data File : VX014626.D
 Acq On : 15 Jan 2020 18:10
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Jan 16 03:14:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jan 15 12:41:12 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	30.000	30.000	0.0	96	0.00
2 M	Dichlorodifluoromethane	20.000	20.517	-2.6	98	0.00
3 M	Chloromethane	20.000	20.700	-3.5	100	0.00
4 M	Vinyl Chloride	20.000	20.134	-0.7	98	0.00
5 M	Bromomethane	20.000	18.916	5.4	110	0.00
6 M	Chloroethane	20.000	19.950	0.3	95	0.02
7 M	Trichlorofluoromethane	20.000	19.881	0.6	101	0.01
8 T	Diethyl Ether	20.000	20.170	-0.9	103	0.00
9	1,1,2-Trichlorotrifluoroeth	20.000	20.003	-0.0	100	0.01
10 M	1,1-Dichloroethene	20.000	19.846	0.8	101	0.00
11	Methyl Iodide	20.000	17.910	10.4	94	0.00
12	Methyl Acetate	20.000	20.583	-2.9	101	0.00
13 M	Acrolein	100.000	89.580	10.4	90	0.00
14 M	Acrylonitrile	100.000	100.459	-0.5	99	0.00
15 M	Acetone	100.000	87.180	12.8	82	0.00
16 M	Carbon Disulfide	20.000	19.204	4.0	99	0.00
17	Allyl chloride	20.000	19.296	3.5	99	0.00
18 M	Methylene Chloride	20.000	20.018	-0.1	100	0.00
19 M	trans-1,2-Dichloroethene	20.000	19.656	1.7	100	0.00
20 T	Diisopropyl ether	20.000	20.193	-1.0	101	0.00
21 M	1,1-Dichloroethane	20.000	19.553	2.2	99	0.00
22 M	cis-1,2-Dichloroethene	20.000	19.490	2.6	100	0.00
23 M	tert-Butyl Alcohol	100.000	100.377	-0.4	104	-0.02
24 M	Methyl tert-Butyl Ether	20.000	19.865	0.7	100	0.00
25 M	Chloroform	20.000	19.458	2.7	99	0.00
26	Cyclohexane	20.000	19.768	1.2	101	0.00
27 s	1,2-Dichloroethane-d4	30.000	29.802	0.7	96	0.00
28 I	1,4-Difluorobenzene	30.000	30.000	0.0	97	0.00
29	1,1-Dichloropropene	20.000	19.734	1.3	101	0.00
30 M	2-Butanone	100.000	94.772	5.2	93	0.00
31	2,2-Dichloropropane	20.000	19.434	2.8	100	0.00
32 M	1,1,1-Trichloroethane	20.000	20.137	-0.7	104	0.00
33 M	Carbon Tetrachloride	20.000	19.531	2.3	101	0.00
34 M	Benzene	20.000	20.056	-0.3	100	0.00
35	Methacrylonitrile	20.000	19.412	2.9	97	0.00
36 M	1,2-Dichloroethane	20.000	20.042	-0.2	100	0.00
37 M	Trichloroethene	20.000	20.179	-0.9	101	0.00
38	Methylcyclohexane	20.000	19.556	2.2	99	0.00
39 M	1,2-Dichloropropane	20.000	19.867	0.7	100	0.00
40	Dibromomethane	20.000	19.612	1.9	100	0.00
41 M	Bromodichloromethane	20.000	19.202	4.0	100	0.00
42 M	Vinyl Acetate	100.000	101.070	-1.1	102	0.00
43	Ethyl Acetate	20.000	19.806	1.0	97	0.00
44	Isopropyl Acetate	20.000	19.872	0.6	101	0.00
45 T	1,4-Dioxane	400.000	426.418	-6.6	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX011520\
 Data File : VX014626.D
 Acq On : 15 Jan 2020 18:10
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Jan 16 03:14:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jan 15 12:41:12 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
46	Methyl methacrylate	20.000	19.793	1.0	101	0.00
47	n-amyl Acetate	20.000	19.242	3.8	104	0.00
48 M	t-1,3-Dichloropropene	20.000	18.986	5.1	99	0.00
49 T	cis-1,3-Dichloropropene	20.000	19.242	3.8	100	0.00
50 M	1,1,2-Trichloroethane	20.000	19.905	0.5	99	0.00
51	Ethyl methacrylate	20.000	19.259	3.7	103	0.00
52	1,3-Dichloropropane	20.000	20.147	-0.7	102	0.00
53 M	Dibromochloromethane	20.000	19.334	3.3	102	0.00
54 M	1,2-Dibromoethane	20.000	20.169	-0.8	101	0.00
55 M	2-Chloroethyl vinyl ether	100.000	94.328	5.7	97	0.00
56 M	Bromoform	20.000	18.123	9.4	95	0.00
57 I	Chlorobenzene-d5	30.000	30.000	0.0	99	0.00
58 M	4-Methyl-2-Pentanone	100.000	100.518	-0.5	101	0.00
59 M	2-Hexanone	100.000	97.456	2.5	97	0.00
60 S	4-Bromofluorobenzene	30.000	29.636	1.2	98	0.00
61 M	Tetrachloroethene	20.000	18.896	5.5	95	0.00
62 M	Toluene	20.000	19.973	0.1	101	0.00
63 S	Toluene-d8	30.000	30.376	-1.3	99	0.00
64 M	Chlorobenzene	20.000	19.868	0.7	101	0.00
65	1,1,1,2-Tetrachloroethane	20.000	19.313	3.4	98	0.00
66 M	Ethyl Benzene	20.000	19.938	0.3	103	0.00
67 M	m/p-Xylenes	40.000	39.033	2.4	100	0.00
68 M	o-Xylene	20.000	19.147	4.3	99	0.00
69 M	Styrene	20.000	19.490	2.6	102	0.00
70	Isopropylbenzene	20.000	19.684	1.6	102	0.00
71 M	1,1,2,2-Tetrachloroethane	20.000	19.441	2.8	99	0.00
72	1,2,3-Trichloropropane	20.000	19.724	1.4	101	0.00
73	Bromobenzene	20.000	19.312	3.4	100	0.00
74	n-propylbenzene	20.000	19.411	2.9	102	0.00
75	2-Chlorotoluene	20.000	19.434	2.8	101	0.00
76	1,3,5-Trimethylbenzene	20.000	19.909	0.5	103	0.00
77	t-1,4-Dichloro-2-butene	20.000	17.546	12.3	98	0.00
78	4-Chlorotoluene	20.000	19.378	3.1	101	0.00
79	tert-butylbenzene	20.000	19.841	0.8	104	0.00
80	1,2,4-Trimethylbenzene	20.000	19.782	1.1	102	0.00
81	sec-Butylbenzene	20.000	19.598	2.0	103	0.00
82	p-Isopropyltoluene	20.000	19.405	3.0	101	0.00
83 M	1,3-Dichlorobenzene	20.000	19.349	3.3	101	0.00
84 M	1,4-Dichlorobenzene	20.000	19.029	4.9	102	0.00
85	n-Butylbenzene	20.000	18.561	7.2	102	0.00
86 T	Hexachloroethane	20.000	18.516	7.4	101	0.00
87 M	1,2-Dichlorobenzene	20.000	18.976	5.1	99	0.00
88	1,2-Dibromo-3-Chloropropane	20.000	19.479	2.6	102	0.00
89	1,2,4-Trichlorobenzene	20.000	18.725	6.4	101	0.00
90	Hexachlorobutadiene	20.000	19.106	4.5	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX011520\
Data File : VX014626.D
Acq On : 15 Jan 2020 18:10
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC020

Quant Time: Jan 16 03:14:53 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jan 15 12:41:12 2020
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
91 M	Naphthalene	20.000	19.524	2.4	103	0.00
92	1,2,3-Trichlorobenzene	20.000	19.299	3.5	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0