

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011521\
 Data File : VX020639.D
 Acq On : 15 Jan 2021 18:32
 Operator : JC/MD
 Sample : M1158-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Jan 16 00:57:29 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	195317	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	346801	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	325378	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	139469	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	138659	59.10	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	118.20%
35) Dibromofluoromethane	5.464	113	118337	49.25	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.50%
50) Toluene-d8	8.695	98	414873	48.25	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.50%
62) 4-Bromofluorobenzene	11.122	95	143321	47.20	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.40%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011521\
Data File : VX020639.D
Acq On : 15 Jan 2021 18:32
Operator : JC/MD
Sample : M1158-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Jan 16 00:57:29 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

