

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011521\
 Data File : VX020646.D
 Acq On : 15 Jan 2021 21:13
 Operator : JC/MD
 Sample : M1157-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Jan 16 00:59:54 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	189792	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	335769	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	321498	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	142831	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	136265	59.77	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	119.54%
35) Dibromofluoromethane	5.465	113	114952	49.41	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.82%
50) Toluene-d8	8.696	98	412718	49.58	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.16%
62) 4-Bromofluorobenzene	11.122	95	147532	50.18	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.36%
Target Compounds						
95) Naphthalene	13.817	128	21809	5.75	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011521\
Data File : VX020646.D
Acq On : 15 Jan 2021 21:13
Operator : JC/MD
Sample : M1157-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Jan 16 00:59:54 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

