

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
 Data File : VX026478.D
 Acq On : 15 Jan 2022 01:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 17 00:43:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	106422	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	177146	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	155095	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73722	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71980	48.418	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.840%
35) Dibromofluoromethane	5.391	113	54692	47.810	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.620%
50) Toluene-d8	8.652	98	200776	48.705	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.400%
62) 4-Bromofluorobenzene	11.085	95	75190	51.345	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.700%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.172	85	58114	46.245	ug/l	99
3) Chloromethane	1.294	50	55154	47.876	ug/l	99
4) Vinyl Chloride	1.379	62	58501	48.367	ug/l	100
5) Bromomethane	1.617	94	25483	45.672	ug/l	97
6) Chloroethane	1.690	64	27521	44.997	ug/l	99
7) Trichlorofluoromethane	1.892	101	89480	46.858	ug/l	99
8) Diethyl Ether	2.135	74	33258	49.218	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.337	101	55968	48.094	ug/l	99
10) Methyl Iodide	2.458	142	67683	53.669	ug/l	99
11) Tert butyl alcohol	3.019	59	69690	336.980	ug/l #	75
12) 1,1-Dichloroethene	2.324	96	54090	48.667	ug/l	97
13) Acrolein	2.245	56	56424	256.759	ug/l	99
14) Allyl chloride	2.672	41	102598	50.196	ug/l	99
15) Acrylonitrile	3.068	53	166609	261.712	ug/l	100
16) Acetone	2.391	43	149083	196.647	ug/l	98
17) Carbon Disulfide	2.519	76	147893	45.577	ug/l	100
18) Methyl Acetate	2.708	43	78212	51.893	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	200454	49.706	ug/l	100
20) Methylene Chloride	2.794	84	59643	47.616	ug/l	99
21) trans-1,2-Dichloroethene	3.099	96	58996	49.032	ug/l	97
22) Diisopropyl ether	3.763	45	206199	49.665	ug/l	92
23) Vinyl Acetate	3.727	43	893059	261.314	ug/l	99
24) 1,1-Dichloroethane	3.617	63	111234	49.609	ug/l	99
25) 2-Butanone	4.568	43	239170	242.585	ug/l	99
26) 2,2-Dichloropropane	4.483	77	76605	46.091	ug/l	100
27) cis-1,2-Dichloroethene	4.495	96	67810	49.380	ug/l	100
28) Bromochloromethane	4.903	49	46585	50.598	ug/l	100
29) Tetrahydrofuran	5.013	42	157598	269.142	ug/l	100
30) Chloroform	5.104	83	115420	49.542	ug/l	98
31) Cyclohexane	5.476	56	103165	47.048	ug/l	98
32) 1,1,1-Trichloroethane	5.391	97	102448	49.993	ug/l	99
36) 1,1-Dichloropropene	5.702	75	85586	47.393	ug/l	100
37) Ethyl Acetate	4.720	43	95393	52.219	ug/l	99
38) Carbon Tetrachloride	5.684	117	90142	48.527	ug/l	99
39) Methylcyclohexane	7.384	83	101901	46.781	ug/l	99
40) Benzene	6.043	78	240918	48.266	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
 Data File : VX026478.D
 Acq On : 15 Jan 2022 01:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 17 00:43:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	54086	52.265	ug/l	99
42) 1,2-Dichloroethane	6.092	62	94206	48.238	ug/l	100
43) Isopropyl Acetate	6.348	43	152616	51.718	ug/l	99
44) Trichloroethene	7.128	130	69839	49.308	ug/l	96
45) 1,2-Dichloropropane	7.433	63	62139	49.873	ug/l	98
46) Dibromomethane	7.586	93	43085	49.054	ug/l	99
47) Bromodichloromethane	7.823	83	87425	49.264	ug/l	99
48) Methyl methacrylate	7.695	41	75784	51.821	ug/l	99
49) 1,4-Dioxane	7.695	88	27129	1130.413	ug/l	98
51) 4-Methyl-2-Pentanone	8.579	43	466159	266.333	ug/l	99
52) Toluene	8.720	92	150893	49.296	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	87155	46.095	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	97134	50.961	ug/l	98
55) 1,1,2-Trichloroethane	9.152	97	59294	49.639	ug/l	98
56) Ethyl methacrylate	9.122	69	99058	54.113	ug/l	99
57) 1,3-Dichloropropane	9.311	76	104370	50.398	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	228787	238.348	ug/l	99
59) 2-Hexanone	9.433	43	349519	260.908	ug/l	100
60) Dibromochloromethane	9.524	129	65136	52.328	ug/l	99
61) 1,2-Dibromoethane	9.610	107	63542	51.051	ug/l	99
64) Tetrachloroethene	9.274	164	70652	48.223	ug/l	97
65) Chlorobenzene	10.079	112	156465	47.906	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.164	131	59892	49.352	ug/l	99
67) Ethyl Benzene	10.195	91	288799	48.659	ug/l	100
68) m/p-Xylenes	10.305	106	219843	98.978	ug/l	99
69) o-Xylene	10.646	106	105816	49.372	ug/l	100
70) Styrene	10.658	104	178575	51.062	ug/l	99
71) Bromoform	10.798	173	43279	46.746	ug/l #	99
73) Isopropylbenzene	10.963	105	282399	46.732	ug/l	100
74) N-amyl acetate	10.847	43	123338	52.869	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	79197	48.827	ug/l	100
76) 1,2,3-Trichloropropane	11.243	75	97249	59.912	ug/l	91
77) Bromobenzene	11.201	156	65540	47.343	ug/l	99
78) n-propylbenzene	11.304	91	324760	47.128	ug/l	100
79) 2-Chlorotoluene	11.365	91	194528	46.552	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	237144	47.601	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	23968	46.919	ug/l	95
82) 4-Chlorotoluene	11.457	91	225165	46.884	ug/l	100
83) tert-Butylbenzene	11.719	119	220788	47.474	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	236723	47.507	ug/l	99
85) sec-Butylbenzene	11.890	105	289062	48.052	ug/l	100
86) p-Isopropyltoluene	12.012	119	244632	48.471	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	122064	47.040	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	121913	46.417	ug/l	99
89) n-Butylbenzene	12.335	91	208417	48.121	ug/l	100
90) Hexachloroethane	12.542	117	39069	46.545	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	118914	48.493	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.944	75	20656	52.692	ug/l	99
93) 1,2,4-Trichlorobenzene	13.591	180	75478	48.009	ug/l	98
94) Hexachlorobutadiene	13.725	225	30771	46.365	ug/l	100
95) Naphthalene	13.780	128	268110	51.127	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	76452	49.447	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
Data File : VX026478.D
Acq On : 15 Jan 2022 01:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jan 17 00:43:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
 Data File : VX026478.D
 Acq On : 15 Jan 2022 01:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :


Quant Time: Jan 17 00:43:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 106422
Ion Ratio Lower Upper
168 100
99 64.7 50.3 75.5



#2
Dichlorodifluoromethane
Concen: 46.245 ug/l
RT: 1.172 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 85 Resp: 58114
Ion Ratio Lower Upper
85 100
87 32.2 16.4 49.1









#3
Chloromethane
Concen: 47.876 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 55154
Ion Ratio Lower Upper
50 100
52 33.3 26.1 39.1



#4
Vinyl Chloride
Concen: 48.367 ug/l
RT: 1.379 min Scan# 48
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 62 Resp: 58501
Ion Ratio Lower Upper
62 100
64 32.7 26.2 39.2









#5
Bromomethane
Concen: 45.672 ug/l
RT: 1.617 min Scan# 87
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 25483
Ion Ratio Lower Upper
94 100
96 97.5 75.9 113.9



#6
Chloroethane
Concen: 44.997 ug/l
RT: 1.690 min Scan# 99
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14


Tgt Ion: 64 Resp: 27521
Ion Ratio Lower Upper
64 100
66 31.3 25.4 38.2








#7
 Trichlorofluoromethane
 Concen: 46.858 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14

Instrument :
 MSVOA_X
 ClientSampleId :






Tgt Ion:101 Resp: 89480
 Ion Ratio Lower Upper
 101 100
 103 64.6 51.3 76.9



#8
 Diethyl Ether
 Concen: 49.218 ug/l
 RT: 2.135 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14

Tgt Ion: 74 Resp: 33258
 Ion Ratio Lower Upper
 74 100
 45 101.0 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.094 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 55968
Ion Ratio Lower Upper
101 100
85 45.5 35.3 52.9
151 77.4 61.4 92.2



#10
Methyl Iodide
Concen: 53.669 ug/l
RT: 2.458 min Scan# 225
Delta R.T. -0.001 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion:142 Resp: 67683
Ion Ratio Lower Upper
142 100
127 47.0 38.1 57.1
141 14.4 11.2 16.8





#11

Tert butyl alcohol

Concen: 336.980 ug/l

RT: 3.019 min Scan# 315

Delta R.T. 0.012 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 69690

Ion Ratio Lower Upper

59 100

57 0.0 7.2 10.8#



#12

1,1-Dichloroethene

Concen: 48.667 ug/l

RT: 2.324 min Scan# 203

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14




Tgt Ion: 96 Resp: 54090


Ion Ratio Lower Upper

96 100

61 167.7 138.5 207.7


98 62.9 51.8 77.6








#13
Acrolein
Concen: 256.759 ug/l
RT: 2.245 min Scan# 19
Delta R.T. 0.006 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 56424
Ion Ratio Lower Upper
56 100
55 71.5 56.9 85.3



#14
Allyl chloride
Concen: 50.196 ug/l
RT: 2.672 min Scan# 260
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14


Tgt Ion: 41 Resp: 102598
Ion Ratio Lower Upper
41 100
39 66.8 54.2 81.4
76 30.7 25.2 37.8








#15
Acrylonitrile
Concen: 261.712 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 166609
Ion Ratio Lower Upper
53 100
52 83.2 66.6 100.0
51 37.0 29.5 44.3



#16
Acetone
Concen: 196.647 ug/l
RT: 2.391 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 43 Resp: 149083
Ion Ratio Lower Upper
43 100
58 29.9 23.0 34.6









#17
Carbon Disulfide
Concen: 45.577 ug/l
RT: 2.519 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 76 Resp: 147893
Ion Ratio Lower Upper
76 100
78 9.4 7.5 11.3



#18
Methyl Acetate
Concen: 51.893 ug/l
RT: 2.708 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 43 Resp: 78212
Ion Ratio Lower Upper
43 100
74 22.5 18.3 27.5









#19
Methyl tert-butyl Ether
Concen: 49.706 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 73 Resp: 200454
Ion Ratio Lower Upper
73 100
57 21.9 17.4 26.2



#20
Methylene Chloride
Concen: 47.616 ug/l
RT: 2.794 min Scan# 280
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 84 Resp: 59643
Ion Ratio Lower Upper
84 100
49 125.9 101.5 152.3
51 41.1 31.2 46.8
86 65.3 51.3 76.9









#21
trans-1,2-Dichloroethene
Concen: 49.032 ug/l
RT: 3.099 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 58996
Ion Ratio Lower Upper
96 100
61 144.3 112.9 169.3
98 64.4 49.5 74.3



#22
Diisopropyl ether
Concen: 49.665 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 45 Resp: 206199
Ion Ratio Lower Upper
45 100
43 84.6 60.6 91.0
87 26.1 22.0 33.0
59 10.2 9.0 13.4









#23
 Vinyl Acetate
 Concen: 261.314 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 893059
 Ion Ratio Lower Upper
 43 100
 86 10.3 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 49.609 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. -0.000 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14

Tgt Ion: 63 Resp: 111234
 Ion Ratio Lower Upper
 63 100
 98 7.2 3.7 11.1
 100 4.6 2.4 7.2









#25
2-Butanone
Concen: 242.585 ug/l
RT: 4.568 min Scan# 51
Delta R.T. 0.006 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 239170
Ion Ratio Lower Upper
43 100
72 25.2 20.4 30.6



#26
2,2-Dichloropropane
Concen: 46.091 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 77 Resp: 76605
Ion Ratio Lower Upper
77 100
97 22.0 11.0 33.0





#27

cis-1,2-Dichloroethene

Concen: 49.380 ug/l

RT: 4.495 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 67810

Ion Ratio Lower Upper

96 100

61 147.9 0.0 296.6

98 64.7 0.0 128.8



#28

Bromochloromethane

Concen: 50.598 ug/l

RT: 4.903 min Scan# 626

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14




Tgt Ion: 49 Resp: 46585


Ion Ratio Lower Upper

49 100

129 2.2 0.0 4.2

130 78.2 62.4 93.6





#29

Tetrahydrofuran

Concen: 269.142 ug/l

RT: 5.013 min Scan# 644

Delta R.T. -0.000 min


Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 42 Resp: 157598

Ion Ratio Lower Upper

42 100

72 43.1 34.6 52.0

71 39.9 31.8 47.8



#30

Chloroform

Concen: 49.542 ug/l

RT: 5.104 min Scan# 659

Delta R.T. 0.006 min

Lab File: VX026478.D




Acq: 15 Jan 2022 01:14


Tgt Ion: 83 Resp: 115420

Ion Ratio Lower Upper

83 100


85 66.3 51.6 77.4








#31
Cyclohexane
Concen: 47.048 ug/l
RT: 5.476 min Scan# 720
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 103165
Ion Ratio Lower Upper
56 100
69 30.1 25.8 38.6
84 84.3 68.9 103.3



#32
1,1,1-Trichloroethane
Concen: 49.993 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 97 Resp: 102448
Ion Ratio Lower Upper
97 100
99 64.6 51.3 76.9
61 44.3 35.0 52.6





#33

1,2-Dichloroethane-d4

Concen: 48.418 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 71980

Ion Ratio Lower Upper

65 100

67 50.8 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.006 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14




Tgt Ion: 114 Resp: 177146

Ion Ratio Lower Upper

114 100

63 21.0 0.0 42.8

88 16.4 0.0 33.0





#35

Dibromofluoromethane

Concen: 47.810 ug/l

RT: 5.391 min Scan# 706

Delta R.T. -0.000 min


Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:113 Resp: 54692

Ion Ratio Lower Upper

113 100

111 103.8 81.2 121.8

192 19.1 15.0 22.4



#36

1,1-Dichloropropene

Concen: 47.393 ug/l

RT: 5.702 min Scan# 757

Delta R.T. 0.006 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14




Tgt Ion: 75 Resp: 85586


Ion Ratio Lower Upper

75 100

110 34.2 17.2 51.6

77 30.8 24.6 36.8





#37

Ethyl Acetate

Concen: 52.219 ug/l

RT: 4.720 min Scan# 596

Delta R.T. -0.000 min


Lab File: VX026478.D

Acq: 15 Jan 2022 01:14



Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:	43	Resp:	95393
Ion	Ratio	Lower	Upper
43	100		
61	13.1	10.6	16.0
70	10.9	8.5	12.7



#38

Carbon Tetrachloride

Concen: 48.527 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14


Tgt Ion:	117	Resp:	90142
Ion	Ratio	Lower	Upper
117	100		
119	93.7	74.7	112.1
121	30.4	25.4	38.0








#39
Methylcyclohexane
Concen: 46.781 ug/l
RT: 7.384 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 83 Resp: 101901
Ion Ratio Lower Upper
83 100
55 80.8 63.6 95.4
98 48.3 38.4 57.6



#40
Benzene
Concen: 48.266 ug/l
RT: 6.043 min Scan# 813
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14


Tgt Ion: 78 Resp: 240918
Ion Ratio Lower Upper
78 100
77 23.4 18.9 28.3








#41
Methacrylonitrile
Concen: 52.265 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41	Resp: 54086
Ion Ratio	Lower Upper
41 100	
39 55.0	45.4 68.0
67 66.2	53.2 79.8
52 28.4	23.0 34.6



#42
1,2-Dichloroethane
Concen: 48.238 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 62	Resp: 94206
Ion Ratio	Lower Upper
62 100	
98 8.7	0.0 17.8





#43

Isopropyl Acetate

Concen: 51.718 ug/l

RT: 6.348 min Scan# 862

Delta R.T. 0.006 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 43 Resp: 152616

Ion Ratio Lower Upper

43 100

61 19.1 15.4 23.0

87 13.1 10.2 15.2



#44

Trichloroethene

Concen: 49.308 ug/l

RT: 7.128 min Scan# 991

Delta R.T. -0.000 min

Lab File: VX026478.D




Acq: 15 Jan 2022 01:14

Tgt Ion:130 Resp: 69839

Ion Ratio Lower Upper

130 100

95 97.2 0.0 202.6





#45

1,2-Dichloropropane

Concen: 49.873 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 62139

Ion Ratio Lower Upper

63 100

65 30.9 23.8 35.6



#46

Dibromomethane

Concen: 49.054 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14




Tgt Ion: 93 Resp: 43085


Ion Ratio Lower Upper

93 100

95 83.3 66.1 99.1

174 99.2 78.7 118.1









#47
 Bromodichloromethane
 Concen: 49.264 ug/l
 RT: 7.823 min Scan# 1105
 Delta R.T. -0.000 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 83 Resp: 87425
 Ion Ratio Lower Upper
 83 100
 85 65.1 51.4 77.0
 127 8.5 6.9 10.3



#48
 Methyl methacrylate
 Concen: 51.821 ug/l
 RT: 7.695 min Scan# 1084
 Delta R.T. -0.000 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14

Tgt Ion: 41 Resp: 75784
 Ion Ratio Lower Upper
 41 100
 69 76.9 62.2 93.4
 39 56.1 45.4 68.0





#49

1,4-Dioxane

Concen: 1130.413 ug/l

RT: 7.695 min Scan# 1084

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 27129

Ion Ratio Lower Upper

88 100

43 35.9 29.9 44.9

58 73.6 57.8 86.6



#50

Toluene-d8

Concen: 48.705 ug/l

RT: 8.652 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX026478.D




Acq: 15 Jan 2022 01:14


Tgt Ion: 98 Resp: 200776

Ion Ratio Lower Upper

98 100

100 66.6 53.2 79.8









#51
4-Methyl-2-Pentanone
Concen: 266.333 ug/l
RT: 8.579 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 466159
Ion Ratio Lower Upper
43 100
58 38.4 30.5 45.7



#52
Toluene
Concen: 49.296 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 92 Resp: 150893
Ion Ratio Lower Upper
92 100
91 170.1 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 46.095 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 87155

Ion Ratio Lower Upper

75 100

77 31.2 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 50.961 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14




Tgt Ion: 75 Resp: 97134


Ion Ratio Lower Upper

75 100

77 32.2 26.2 39.2


39 52.9 43.5 65.3








#55
1,1,2-Trichloroethane
Concen: 49.639 ug/l
RT: 9.152 min Scan# 1318
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:	97	Resp:	59294
Ion	Ratio	Lower	Upper
97	100		
83	86.6	68.4	102.6
85	53.5	44.0	66.0
99	63.9	49.8	74.6



#56
Ethyl methacrylate
Concen: 54.113 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.006 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion:	69	Resp:	99058
Ion	Ratio	Lower	Upper
69	100		
41	70.5	57.2	85.8
39	37.7	30.9	46.3





#57

1,3-Dichloropropane

Concen: 50.398 ug/l

RT: 9.311 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 104370

Ion Ratio Lower Upper

76 100

78 32.3 25.7 38.5



#58

2-Chloroethyl Vinyl ether

Concen: 238.348 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX026478.D




Acq: 15 Jan 2022 01:14


Tgt Ion: 63 Resp: 228787

Ion Ratio Lower Upper

63 100

106 29.3 23.0 34.6





#59

2-Hexanone

Concen: 260.908 ug/l

RT: 9.433 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 349519

Ion Ratio Lower Upper

43 100

58 53.0 26.4 79.2



#60

Dibromochloromethane

Concen: 52.328 ug/l

RT: 9.524 min Scan# 1384

Delta R.T. -0.000 min

Lab File: VX026478.D




Acq: 15 Jan 2022 01:14

Tgt Ion: 129 Resp: 65136

Ion Ratio Lower Upper

129 100

127 77.5 38.5 115.5









#61
1,2-Dibromoethane
Concen: 51.051 ug/l
RT: 9.610 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 63542
Ion Ratio Lower Upper
107 100
109 93.6 75.5 113.3



#62
4-Bromofluorobenzene
Concen: 51.345 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.006 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14


Tgt Ion: 95 Resp: 75190
Ion Ratio Lower Upper
95 100
174 73.5 0.0 148.6
176 70.2 0.0 141.4







#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:117 Resp: 155095
Ion Ratio Lower Upper
117 100
82 58.3 45.2 67.8
119 32.5 26.2 39.2




#64
Tetrachloroethene
Concen: 48.223 ug/l
RT: 9.274 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14



Tgt Ion:164 Resp: 70652
Ion Ratio Lower Upper
164 100
166 128.4 101.8 152.6
129 99.1 83.3 124.9
131 95.2 78.0 117.0









#65
Chlorobenzene
Concen: 47.906 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 156465
Ion Ratio Lower Upper
112 100
114 32.3 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 49.352 ug/l
RT: 10.164 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion:131 Resp: 59892
Ion Ratio Lower Upper
131 100
133 94.4 47.9 143.8
119 62.5 31.7 95.1









#67
Ethyl Benzene
Concen: 48.659 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 288799
Ion Ratio Lower Upper
91 100
106 30.7 24.7 37.1



#68
m/p-Xylenes
Concen: 98.978 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 106 Resp: 219843
Ion Ratio Lower Upper
106 100
91 205.5 165.3 247.9









#69
o-Xylene
Concen: 49.372 ug/l
RT: 10.646 min Scan# 1567
Delta R.T. 0.006 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 105816
Ion Ratio Lower Upper
106 100
91 219.7 109.7 329.0



#70
Styrene
Concen: 51.062 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion:104 Resp: 178575
Ion Ratio Lower Upper
104 100
78 52.3 42.5 63.7
103 54.1 43.3 64.9





#71

Bromoform

Concen: 46.746 ug/l

RT: 10.798 min Scan# 1593

Delta R.T. -0.001 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:173 Resp: 43279

Ion Ratio Lower Upper

173 100

175 47.7 24.3 72.9

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14




Tgt Ion:152 Resp: 73722


Ion Ratio Lower Upper

152 100

115 86.6 44.2 132.6

150 173.0 0.0 348.2





#73

Isopropylbenzene

Concen: 46.732 ug/l

RT: 10.963 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 282399

Ion Ratio Lower Upper

105 100

120 26.0 13.1 39.1



#74

N-ethyl acetate

Concen: 52.869 ug/l

RT: 10.847 min Scan# 1601

Delta R.T. 0.006 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Tgt Ion: 43 Resp: 123338




Ion Ratio Lower Upper


43 100

70 41.6 32.9 49.3

55 23.5 18.3 27.5

61 23.2 18.9 28.3





#75

1,1,2,2-Tetrachloroethane

Concen: 48.827 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 79197

Ion Ratio Lower Upper

83 100

131 10.7 5.3 15.9

85 63.8 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 59.912 ug/l

RT: 11.243 min Scan# 1666

Delta R.T. -0.000 min

Lab File: VX026478.D




Acq: 15 Jan 2022 01:14


Tgt Ion: 75 Resp: 97249

Ion Ratio Lower Upper

75 100


77 31.2 18.3 54.8








#77
Bromobenzene
Concen: 47.343 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:156 Resp: 65540
Ion Ratio Lower Upper
156 100
77 167.9 84.5 253.4
158 95.7 48.1 144.4




#78
n-propylbenzene
Concen: 47.128 ug/l
RT: 11.304 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Tgt Ion: 91 Resp: 324760
Ion Ratio Lower Upper
91 100
120 23.0 11.6 34.7








#79
2-Chlorotoluene
Concen: 46.552 ug/l
RT: 11.365 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 91 Resp: 194528
Ion Ratio Lower Upper
91 100
126 33.5 16.4 49.2



#80
1,3,5-Trimethylbenzene
Concen: 47.601 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion: 105 Resp: 237144
Ion Ratio Lower Upper
105 100
120 47.8 24.0 72.0









#81
trans-1,4-Dichloro-2-butene
Concen: 46.919 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 23968
Ion Ratio Lower Upper
75 100
53 111.3 94.0 141.0
89 45.0 37.0 55.4



#82
4-Chlorotoluene
Concen: 46.884 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14


Tgt Ion: 91 Resp: 225165
Ion Ratio Lower Upper
91 100
126 28.5 14.1 42.4








#83
 tert-Butylbenzene
 Concen: 47.474 ug/l
 RT: 11.719 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14

Instrument :
 MSVOA_X
 ClientSampleId :



Tgt Ion:119 Resp: 220788




Ion	Ratio	Lower	Upper
119	100		
91	63.2	31.4	94.0
134	23.8	11.8	35.3




#84
 1,2,4-Trimethylbenzene
 Concen: 47.507 ug/l
 RT: 11.756 min Scan# 1750
 Delta R.T. 0.006 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14

Tgt Ion:105 Resp: 236723

Ion	Ratio	Lower	Upper
105	100		
120	44.5	21.9	65.5









#85
 sec-Butylbenzene
 Concen: 48.052 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14




Instrument :
 MSVOA_X
 ClientSampleId :

Tgt Ion:105 Resp: 289062
 Ion Ratio Lower Upper
 105 100
 134 20.1 10.1 30.1



#86
 p-Isopropyltoluene
 Concen: 48.471 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14

Tgt Ion:119 Resp: 244632
 Ion Ratio Lower Upper
 119 100
 134 25.7 13.0 39.0
 91 25.0 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 47.040 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 122064

Ion Ratio Lower Upper

146 100

111 41.9 21.3 63.7

148 64.0 31.6 95.0



#88

1,4-Dichlorobenzene

Concen: 46.417 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14




Tgt Ion:146 Resp: 121913


Ion Ratio Lower Upper

146 100

111 41.2 20.2 60.6

148 63.9 31.4 94.0





#89

n-Butylbenzene

Concen: 48.121 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. -0.000 min

Lab File: VX026478.D

Acq: 15 Jan 2022 01:14

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 208417

Ion Ratio Lower Upper

91 100

92 54.6 27.2 81.6

134 26.3 13.1 39.1



#90

Hexachloroethane

Concen: 46.545 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX026478.D




Acq: 15 Jan 2022 01:14


Tgt Ion: 117 Resp: 39069

Ion Ratio Lower Upper

117 100


201 71.5 35.6 106.8








#91
 1,2-Dichlorobenzene
 Concen: 48.493 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. -0.000 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14

Instrument :
 MSVOA_X
 ClientSampleId :




Tgt Ion:146 Resp: 118914

Ion	Ratio	Lower	Upper
146	100		
111	43.4	21.8	65.3
148	63.9	32.0	96.2






#92
 1,2-Dibromo-3-Chloropropane
 Concen: 52.692 ug/l
 RT: 12.944 min Scan# 1945
 Delta R.T. -0.000 min
 Lab File: VX026478.D
 Acq: 15 Jan 2022 01:14



Tgt Ion: 75 Resp: 20656


Ion	Ratio	Lower	Upper
75	100		
155	82.0	40.0	119.9
157	103.6	51.6	154.8








#93
1,2,4-Trichlorobenzene
Concen: 48.009 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:180 Resp: 75478
Ion Ratio Lower Upper
180 100
182 95.8 46.7 140.0
145 32.1 16.3 48.8



#94
Hexachlorobutadiene
Concen: 46.365 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14


Tgt Ion:225 Resp: 30771
Ion Ratio Lower Upper
225 100
223 64.0 32.0 96.0
227 64.0 32.3 96.9








#95
Naphthalene
Concen: 51.127 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 268110
Ion Ratio Lower Upper
128 100
127 12.7 10.3 15.5
129 11.3 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 49.447 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX026478.D
Acq: 15 Jan 2022 01:14

Tgt Ion:180 Resp: 76452
Ion Ratio Lower Upper
180 100
182 93.7 47.2 141.6
145 34.3 17.5 52.5

