

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
 Data File : VX026461.D
 Acq On : 14 Jan 2022 18:40
 Operator : JC/MD
 Sample : PB142010TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142010TB

Quant Time: Jan 17 00:41:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	127127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	210687	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189322	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83461	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	86927	48.949	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.900%
35) Dibromofluoromethane	5.391	113	67139	49.347	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.700%
50) Toluene-d8	8.653	98	248479	50.681	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.360%
62) 4-Bromofluorobenzene	11.079	95	93568	53.723	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.440%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	5933	6.551	ug/l	96
20) Methylene Chloride	2.794	84	3204	2.141	ug/l	87
43) Isopropyl Acetate	6.342	43	11557	3.293	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
Data File : VX026461.D
Acq On : 14 Jan 2022 18:40
Operator : JC/MD
Sample : PB142010TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142010TB

Quant Time: Jan 17 00:41:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

