

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
 Data File : VX026463.D
 Acq On : 14 Jan 2022 19:26
 Operator : JC/MD
 Sample : PB142010ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142010ZHE#02

Quant Time: Jan 17 00:41:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	128069	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	215444	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192366	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86310	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	89034	49.767	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.540%
35) Dibromofluoromethane	5.397	113	68803	49.453	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.900%
50) Toluene-d8	8.653	98	253062	50.476	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.960%
62) 4-Bromofluorobenzene	11.085	95	96170	53.998	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.000%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	5780	6.335	ug/l	92
20) Methylene Chloride	2.794	84	3102	2.058	ug/l	98
43) Isopropyl Acetate	6.348	43	11755	3.275	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
Data File : VX026463.D
Acq On : 14 Jan 2022 19:26
Operator : JC/MD
Sample : PB142010ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142010ZHE#02

Quant Time: Jan 17 00:41:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

