

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
 Data File : VX026466.D
 Acq On : 14 Jan 2022 20:36
 Operator : JC/MD
 Sample : PB142010ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB142010ZHE#05

Quant Time: Jan 17 00:41:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128610	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	215625	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83167	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	88646	49.341	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.680%		
35) Dibromofluoromethane	5.397	113	68536	49.220	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.440%		
50) Toluene-d8	8.653	98	252923	50.406	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	100.820%		
62) 4-Bromofluorobenzene	11.085	95	94399	52.959	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	105.920%		
Target Compounds						
16) Acetone	2.386	43	5516	6.021	ug/l	98
20) Methylene Chloride	2.800	84	2897	1.914	ug/l	97
43) Isopropyl Acetate	6.348	43	12126	3.376	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
Data File : VX026466.D
Acq On : 14 Jan 2022 20:36
Operator : JC/MD
Sample : PB142010ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB142010ZHE#05

Quant Time: Jan 17 00:41:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

