

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
 Data File : VX026472.D
 Acq On : 14 Jan 2022 22:55
 Operator : JC/MD
 Sample : N1112-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-4S

Quant Time: Jan 17 00:42:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	145518	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	242953	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96280	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100086	49.236	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.480%
35) Dibromofluoromethane	5.391	113	77450	49.365	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.740%
50) Toluene-d8	8.653	98	283082	50.070	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.140%
62) 4-Bromofluorobenzene	11.079	95	105810	52.684	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.360%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	7384	7.123	ug/l	98
20) Methylene Chloride	2.794	84	3453	2.016	ug/l	95
43) Isopropyl Acetate	6.342	43	10835	2.677	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
Data File : VX026472.D
Acq On : 14 Jan 2022 22:55
Operator : JC/MD
Sample : N1112-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-4S

Quant Time: Jan 17 00:42:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

