

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
 Data File : VX026474.D
 Acq On : 14 Jan 2022 23:41
 Operator : JC/MD
 Sample : N1130-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-5S

Quant Time: Jan 17 00:43:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	138087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234524	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206100	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89053	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96755	50.159	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.320%	
35) Dibromofluoromethane	5.391	113	74789	49.383	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.760%	
50) Toluene-d8	8.653	98	273269	50.072	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.140%	
62) 4-Bromofluorobenzene	11.079	95	99774	51.464	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.920%	
Target Compounds						Qvalue
16) Acetone	2.386	43	275347	279.909	ug/l	99
20) Methylene Chloride	2.794	84	34532	21.247	ug/l	99
43) Isopropyl Acetate	6.348	43	11772	3.013	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
Data File : VX026474.D
Acq On : 14 Jan 2022 23:41
Operator : JC/MD
Sample : N1130-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-5S

Quant Time: Jan 17 00:43:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

