

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
 Data File : VX026477.D
 Acq On : 15 Jan 2022 00:51
 Operator : JC/MD
 Sample : N1136-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-6D

Quant Time: Jan 17 00:43:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	139603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235312	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87766	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	97772	50.136	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.280%		
35) Dibromofluoromethane	5.391	113	75459	49.658	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.320%		
50) Toluene-d8	8.653	98	273376	49.924	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.840%		
62) 4-Bromofluorobenzene	11.079	95	100208	51.514	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 103.020%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	7954	7.998	ug/l	96
20) Methylene Chloride	2.794	84	3601	2.192	ug/l	91
43) Isopropyl Acetate	6.348	43	11720	2.990	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011522\
Data File : VX026477.D
Acq On : 15 Jan 2022 00:51
Operator : JC/MD
Sample : N1136-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6D

Quant Time: Jan 17 00:43:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

