

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011524\
 Data File : VX039749.D
 Acq On : 15 Jan 2024 12:29
 Operator : JC/MD
 Sample : VX0115MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0115MBL01

Quant Time: Jan 16 01:37:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	107201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	201184	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	194206	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	86696	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	81221	45.229	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.460%
35) Dibromofluoromethane	5.379	113	63911	47.165	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.320%
50) Toluene-d8	8.647	98	245603	48.057	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.120%
62) 4-Bromofluorobenzene	11.079	95	95658	47.864	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011524\
Data File : VX039749.D
Acq On : 15 Jan 2024 12:29
Operator : JC/MD
Sample : VX0115MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0115MBL01

Quant Time: Jan 16 01:37:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

