

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011524\
 Data File : VX039754.D
 Acq On : 15 Jan 2024 14:48
 Operator : JC/MD
 Sample : P1145-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-TB-20240110

Quant Time: Jan 16 01:41:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	115881	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	216485	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	203121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90473	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85318	43.952	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.900%
35) Dibromofluoromethane	5.385	113	67606	46.365	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.740%
50) Toluene-d8	8.647	98	261485	47.548	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.100%
62) 4-Bromofluorobenzene	11.079	95	104206	48.456	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011524\
Data File : VX039754.D
Acq On : 15 Jan 2024 14:48
Operator : JC/MD
Sample : P1145-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-TB-20240110

Quant Time: Jan 16 01:41:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

