

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011524\
 Data File : VX039755.D
 Acq On : 15 Jan 2024 15:11
 Operator : JC/MD
 Sample : P1145-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-EB-20240112

Quant Time: Jan 16 01:41:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	104638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	194346	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	183291	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78951	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77064	43.965	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.940%
35) Dibromofluoromethane	5.379	113	61724	47.153	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.300%
50) Toluene-d8	8.647	98	236462	47.896	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.800%
62) 4-Bromofluorobenzene	11.079	95	90827	47.046	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011524\
Data File : VX039755.D
Acq On : 15 Jan 2024 15:11
Operator : JC/MD
Sample : P1145-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-EB-20240112

Quant Time: Jan 16 01:41:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

