

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011524\
 Data File : VX039761.D
 Acq On : 15 Jan 2024 17:27
 Operator : JC/MD
 Sample : P1145-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-200-202

Quant Time: Jan 16 01:43:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	108620	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	201538	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191339	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	84605	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84176	46.262	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.520%
35) Dibromofluoromethane	5.385	113	63739	46.955	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.920%
50) Toluene-d8	8.647	98	246193	48.088	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.180%
62) 4-Bromofluorobenzene	11.079	95	96860	48.381	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.760%
Target Compounds						
16) Acetone	2.380	43	2691	2.630	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011524\
Data File : VX039761.D
Acq On : 15 Jan 2024 17:27
Operator : JC/MD
Sample : P1145-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-200-202

Quant Time: Jan 16 01:43:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

